

Grain Inspection, Packers and Stockyard Admin. (FGIS), USDA

§ 801.7

Item	Tolerance
Air separation	±0.10 percent, mean deviation from standard dockage tester using Hard Red Winter wheat
Riddle separation	±0.10 percent, mean deviation from standard dockage tester using Hard Red Winter wheat
Sieve separation	±0.10 percent, mean deviation from standard dockage tester using Hard Red Winter wheat
Total dockage separation.	±0.15 percent, mean deviation from standard dockage tester using Hard Red Winter wheat

§ 801.5 Tolerance for diverter-type mechanical samplers.

The maintenance tolerance for diverter-type mechanical samplers (pri-

mary, or primary and secondary in combination) used in performing official inspection services shall be ±10 percent, mean deviation from standard sampling device using corn or the same type of grain that the system will be used to sample.

§ 801.6 Tolerances for moisture meters.

(a) The maintenance tolerances for Motomco 919 moisture meters used in performing official inspection services shall be:

(1) Headquarters standard meters:

Moisture range	Tolerance	
	Direct comparison	Sample exchange
Low	±0.05 percent moisture, mean deviation from National standard moisture meter using Hard Red Winter wheat	
Mid	±0.05 percent moisture, mean deviation from National standard moisture meter using Hard Red Winter wheat	
High	±0.05 percent moisture, mean deviation from National standard moisture meter using Hard Red Winter wheat	

(2) All other than Headquarters standard meters:

Moisture range	Tolerance	
	Direct comparison	Sample exchange
Low	±0.15 percent moisture, mean deviation from standard moisture meter using Hard Red Winter wheat	±0.20 percent moisture, mean deviation from standard moisture meter using Hard Red Winter wheat
Mid	±0.10 percent moisture, mean deviation from standard moisture meter using Hard Red Winter wheat	±0.15 percent moisture, mean deviation from standard moisture meter using Hard Red Winter wheat
High	±0.15 percent moisture, mean deviation from standard moisture meter using Hard Red Winter wheat	±0.20 percent moisture, mean deviation from standard moisture meter using Hard Red Winter wheat

(b) The maintenance tolerances for GAC 2100 moisture meters used in performing official inspection services shall be:

(1) Headquarters standard meters. By direct comparison using mid-range Hard Red Winter wheat, ±0.05% mean deviation for the average of the Headquarters standard moisture meters.

(2) All other than Headquarters standard meters. By sample exchange using mid-range Hard Red Winter

wheat, ±0.15% mean deviation from the standard meter.

[63 FR 34554, June 25, 1998]

§ 801.7 Reference methods and tolerances for near-infrared spectroscopy (NIRS) analyzers.

(a) *Reference methods.* (1) The chemical reference protein determinations used to reference and calibrate official NIRS instruments shall be performed in accordance with “Comparison of Kjeldahl Method for Determination of